

# Light Measurement Report

Print date: 26-9-2025

Measurement date and time: 26-9-2025 10:48:27 – Measurement no. VFR-250926-3397-MS

Measurement tracking No. and Link: [VT250926-004750](#)

Operator:



## Laboratory and Equipment

Laboratory Owner and Location  
Goniospectrometer System and Type  
Sensor Name, Calibr. Date and Serial No.  
Spectrometer Manufacturer and Model

Viso Systems, Copenhagen V, Denmark  
LabSpion – Type C, horizontal  
LabSensor Model2 – 11-1-2024 – 3130191315  
Ibsen Photonics, Denmark – Freedom VIS (Custom Viso)

## Measurement Conditions

Number of C-planes and Resolution  
 $\gamma$  (gamma)-Resolution  
Test Distance  
Input Power, Power and Displ. Factors  
Input RMS Voltage and Current  
Frequency of Input Power  
Warm-up Time and Variation

36 planes – 10°  
5°  
2,89 m  
76,8 W – PF 0,97 – DPF 0,97  
230 V – 0,344 A  
50 Hz  
Lamp stabilized in 15 min 2 sec – 2,0%

## Tested Light Source



Product Name  
Item No. and Manufacturer  
Product Description (line 1)  
SWITCH

812713-4000K-80W  
812713-4000K-80W – Dutchfulfillment  
LED HIGHBAY ARGOS | 0-10V | 100W/80W/60W | 60°/90°/120° | CCT

## Main Light Measurement Results

Output – Total Lumen (Up% / Down%)  
Efficiency  
Peak Intensity and Beam Angle  
Correlated Color Temperature, Target/Measured  
Color Rendering Index  
Color Rendering TM30-18  
Color Shift, CIE duv and MacAdam Steps  
Flicker

10760 lm – 0,16% / 99,84%  
140 lm/W  
3602 cd – 114,9°  
CCT = 4000 K / 4128 K  
CRI 71,8  
 $R_f$  73,5 –  $R_g$  95,1  
Duv 0,0020 – SDCM 3,0  
SVM 0 – PstLM 0,1



# Light Measurement Report

Print date: 26-9-2025

Measurement date and time: 26-9-2025 10:48:27 – Measurement no. VFR-250926-3397-MS


Measurement tracking No. and Link: [VT250926-004750](#)

Operator:



## Luminous Intensity diagram

Unit: 0-100% of peak intensity



## Main Values

Output (total Lumen)	10760 lm
Lumen Up% / Down%	0,16% / 99,84%
Peak Intensity	3602 cd
Beam Angle (50%)	114,9°
Beam Angle (90%)	114,9°
Beam Angle (10%)	114,9°

## Cut-off Angle

Average 2,5%	170,9°
--------------	--------

## Field Angle

Average 10%	147,7°
-------------	--------


## Intensity Ratio

In 120° cone	86,9%
In 90° cone	55,6%

**C000-C180**

**C090-C270**

## Linear distribution diagram - Intensity (candela) vs $\gamma$ -angle



# Light Measurement Report

Print date: 26-9-2025


Measurement date and time: 26-9-2025 10:48:27 – Measurement no. VFR-250926-3397-MS

Measurement tracking No. and Link: [VT250926-004750](#)

Operator:




## Iso-intensity Diagram (Iso-candela)



90 %	3235,7 cd
80 %	2876,2 cd
70 %	2516,7 cd
60 %	2157,2 cd
50 %	1797,6 cd
40 %	1438,1 cd
30 %	1078,6 cd
20 %	719,1 cd
10 %	359,5 cd

Peak intensity: 3595,3 cd  
Number of c-planes: 36

## Iso-illuminance Diagram (Iso-lux)



50,0 %	184,2 lx
30,0 %	110,5 lx
10,0 %	36,8 lx
5,0 %	18,4 lx
3,0 %	11,1 lx

Peak illuminance: 368,4 lx  
Mounting height: 3,0 m  
Number of c-planes: 36

# Light Measurement Report

Print date: 26-9-2025

Measurement date and time: 26-9-2025 10:48:27 – Measurement no. VFR-250926-3397-MS

Measurement tracking No. and Link: [VT250926-004750](#)

Operator:



## Color details


Correlated Color Temperature, Target CCT = 4000 K  
 Correlated Color Temperature, Measured CCT = 4128 K  
 Color Rendering Index CRI 71,8  
 Color Rendering Index, R9 (red component) R9 = -27,5  
 Color Rendering TM30-18 R<sub>f</sub> 73,5 – R<sub>g</sub> 95,1  
 Color Quality Scale CQS = 71,9

MacAdam Steps SDCM = 3,0  
 Color coordinates CIE 1931 (x;y) = (0,381;0,377)  
 Color coordinate CIEs 1960 (u;v) = (0,225;0,334)  
 Color deviation from BBL Duv = 0,0020  
 Color coordinate CIEs 1976 (CIELUV) (u';v') = (0,225;0,502)


### CIE 1931



### CIE 1931 – zoomed on Planckian locus




### Color Rendering Index per reference color (CIE 1995)




CRI R values, only R1-R8 are used to calculate final CRI value

R1	R2	R3	R4	R5	R6	R7	R8	R9	R10	R11	R12	R13	R14	R15
69,5	77,2	82,9	72,5	68,7	68,0	81,0	54,7	-27,5	45,4	68,5	40,8	70,4	90,1	62,1

### Spectral power distribution (SPD) / W/nm – 0-100%




### TM30-18 Rf-values per hue bin



TM30 C values, 16 binned values out of total of 99 C values

C1	C2	C3	C4	C5	C6	C7	C8	C9	C10	C11	C12	C13	C14	C15	C16
72,8	73,3	60,3	70,0	78,4	83,7	83,5	80,2	80,6	63,3	63,1	68,7	82,4	78,5	73,0	72,3

### Color Quality Scale by reference color



CQS Q values

Q1	Q2	Q3	Q4	Q5	Q6	Q7	Q8	Q9	Q10	Q11	Q12	Q13	Q14	Q15
74,8	95,2	64,7	61,5	69,6	70,6	73,4	82,8	92,4	76,0	72,0	72,6	74,1	59,4	65,2

# Light Measurement Report

Print date: 26-9-2025


Measurement date and time: 26-9-2025 10:48:27 – Measurement no. VFR-250926-3397-MS

Measurement tracking No. and Link: [VT250926-004750](https://www.viso-systems.com/VT250926-004750)

Operator:



## Beam Details



### Beam intensities from 1 – 20 m

1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	m
3,3	6,6	9,8	13,1	16,4	19,7	23	26,2	29,5	32,8	36,1	39,4	42,7	45,9	49,2	52,5	55,8	59,1	62,3	65,6	ft
3316	829	368	207	133	92	68	52	41	33	27	23	20	17	15	13	11	10	9	8	lux
308	77	34,2	19,3	12,3	8,6	6,3	4,8	3,8	3,1	2,5	2,1	1,8	1,6	1,4	1,2	1,1	1	0,9	0,8	fc

### Intensities in 0° c-plane

0°	5°	10°	15°	20°	25°	30°	35°	40°	45°	50°	55°	60°	65°	70°	75°	80°	85°	90°	95°	γ
3316	3300	3267	3210	3148	3092	3062	3174	3444	3583	3065	2193	1492	935	545	318	186	97	33	6	cd
100%	100%	99%	97%	95%	93%	92%	96%	104%	108%	92%	66%	45%	28%	16%	10%	6%	3%	1%	0%	of 0°val

### Intensities in 90° c-plane

0°	5°	10°	15°	20°	25°	30°	35°	40°	45°	50°	55°	60°	65°	70°	75°	80°	85°	90°	95°	γ
3316	3300	3267	3210	3148	3092	3062	3174	3444	3583	3065	2193	1492	935	545	318	186	97	33	6	cd
100%	100%	99%	97%	95%	93%	92%	96%	104%	108%	92%	66%	45%	28%	16%	10%	6%	3%	1%	0%	of 0°val

### Intensities in 180° c-plane

0°	5°	10°	15°	20°	25°	30°	35°	40°	45°	50°	55°	60°	65°	70°	75°	80°	85°	90°	95°	γ
3316	3300	3267	3210	3148	3092	3062	3174	3444	3583	3065	2193	1492	935	545	318	186	97	33	6	cd
100%	100%	99%	97%	95%	93%	92%	96%	104%	108%	92%	66%	45%	28%	16%	10%	6%	3%	1%	0%	of 0°val

### Intensities in 270° c-plane

0°	5°	10°	15°	20°	25°	30°	35°	40°	45°	50°	55°	60°	65°	70°	75°	80°	85°	90°	95°	γ
3316	3300	3267	3210	3148	3092	3062	3174	3444	3583	3065	2193	1492	935	545	318	186	97	33	6	cd
100%	100%	99%	97%	95%	93%	92%	96%	104%	108%	92%	66%	45%	28%	16%	10%	6%	3%	1%	0%	of 0°val

# Light Measurement Report

Print date: 26-9-2025

Measurement date and time: 26-9-2025 10:48:27 – Measurement no. VFR-250926-3397-MS

Measurement tracking No. and Link: [VT250926-004750](https://vt250926-004750)

Operator:



## Light Planning – UGR table

Uncorrected, comprehensive UGR table according to 117-1995

Reflectances		70	70	50	50	30	70	70	50	50	30
	ρ Ceiling	70	70	50	50	30	70	70	50	50	30
	ρ Walls	50	30	50	30	30	50	30	50	30	30
	ρ Floor	20	20	20	20	20	20	20	20	20	20
Room size		Viewed Crosswise					Viewed Endwise				
H = mounting height above eye level		(Viewing direction orthogonal to lamp length axis)					(Viewing direction parallel to lamp length axis)				
X	Y										
2H	2H	31,1	32,3	31,3	32,6	32,8	31,1	32,3	31,3	32,6	32,8
	3H	31,4	32,7	31,9	32,9	33,1	31,4	32,7	31,9	32,9	33,1
	4H	31,6	32,7	32,0	33,0	33,3	31,6	32,7	32,0	33,0	33,3
	6H	31,7	32,7	32,0	33,0	33,4	31,7	32,7	32,0	33,0	33,4
	8H	31,7	32,7	32,1	33,0	33,4	31,7	32,7	32,1	33,0	33,4
	12H	31,7	32,7	32,1	33,0	33,4	31,7	32,7	32,1	33,0	33,4
4H	2H	31,3	32,4	31,7	32,7	33,0	31,3	32,4	31,7	32,7	33,0
	3H	31,9	32,8	32,3	33,2	33,6	31,9	32,8	32,3	33,2	33,6
	4H	32,0	32,9	32,5	33,3	33,8	32,0	32,9	32,5	33,3	33,8
	6H	32,2	33,0	32,7	33,3	33,7	32,2	33,0	32,7	33,3	33,7
	8H	32,2	33,0	32,7	33,3	33,7	32,2	33,0	32,7	33,3	33,7
	12H	32,2	32,9	32,7	33,3	33,8	32,2	32,9	32,7	33,3	33,8
8H	4H	32,0	32,8	32,6	33,2	33,6	32,0	32,8	32,6	33,2	33,6
	6H	32,3	32,8	32,8	33,3	33,8	32,3	32,8	32,8	33,3	33,8
	8H	32,4	32,9	32,9	33,4	34,0	32,4	32,9	32,9	33,4	34,0
	12H	32,5	32,9	33,1	33,4	34,0	32,5	32,9	33,1	33,4	34,0
12H	4H	32,0	32,6	32,5	33,1	33,5	32,0	32,6	32,5	33,1	33,5
	6H	32,3	32,8	32,8	33,3	33,9	32,3	32,8	32,8	33,3	33,9
	8H	32,4	32,8	33,0	33,3	33,9	32,4	32,8	33,0	33,3	33,9

### Variations with the observer position for the luminaire spacings, S:

S = 1.0H	0,4 / -0,3	0,4 / -0,3
S = 1.5H	1,1 / -1,2	1,1 / -1,2
S = 2.0H	2,1 / -2,3	2,1 / -2,3

## Coefficients of Utilization

Ceiling reflectance	80			70			50			30			10			0		
Wall reflectance	70	50	30	10	70	50	30	10	50	30	10	50	30	10	50	30	10	0
Floor reflectance	20	20	20	20	20	20	20	20	20	20	20	20	20	20	20	20	20	0
RCR	(RCR: Room Cavity Ratio)																	
	Room Values are expressed as percentage of Lumen delivered to the task surface																	
0	119	119	119	119	116	116	116	116	111	111	106	106	106	102	102	102	100	100
1	110	105	101	98	107	103	100	96	99	96	93	95	93	90	91	89	88	86
2	100	93	86	81	98	91	85	80	87	82	78	84	80	76	81	77	74	72
3	92	81	74	67	89	80	73	67	77	71	66	74	69	65	72	67	63	61
4	84	72	64	57	81	71	63	57	68	61	56	66	60	55	64	59	54	52
5	77	64	55	49	75	63	55	49	61	54	48	59	53	47	57	51	47	45
6	71	58	49	42	69	57	48	42	55	47	42	53	46	41	51	46	41	39
7	65	52	43	37	64	51	43	37	49	42	37	48	41	36	47	41	36	34
8	61	47	38	33	59	46	38	33	45	38	32	44	37	32	43	36	32	30
9	56	43	35	29	55	42	34	29	41	34	29	40	33	29	39	33	28	27
10	53	39	31	26	51	39	31	26	38	31	26	37	30	26	36	30	26	24

# Light Measurement Report

Print date: 26-9-2025

Measurement date and time: 26-9-2025 10:48:27 – Measurement no. VFR-250926-3397-MS

Measurement tracking No. and Link: [VT250926-004750](#)


Operator:



## Luminaire budgetary diagram

Uncorrected, comprehensive UGR table according to 117-1995

LAMPS (number of lamps)



### Conditions

H = Room height	Flux = 10760 lm				
H <sub>down</sub> = Lamp distance from ceiling =	0.00 m	Line type	Ceiling reflectance	ρ(%) Wall reflectance	Floor reflectance
H <sub>work</sub> = Work area height from floor =	0.00 m	-----	70	50	30
E <sub>work</sub> = Average lux on work area =	100 lx	—————	50	30	20

### Zonal Lumen Summary

0°-10°	10°-20°	20°-30°	30°-40°	40°-50°	50°-60°	60°-70°	70°-80°	80°-90°
314 lm	908 lm	1431 lm	2008 lm	2710 lm	1979 lm	941 lm	343 lm	108 lm
90°-100°	100°-110°	110°-120°	120°-130°	130°-140°	140°-150°	150°-160°	160°-170°	170°-180°
7,80 lm	0,658 lm	1,45 lm	1,83 lm	1,81 lm	1,61 lm	1,22 lm	0,742 lm	0,262 lm

# Light Measurement Report

Print date: 26-9-2025

Measurement date and time: 26-9-2025 10:48:27 – Measurement no. VFR-250926-3397-MS

Measurement tracking No. and Link: [VT250926-004750](https://www.viso-systems.com/VT250926-004750)

Operator:



## Outdoor Light Planning

### Lumen per Zone

Zone (γ)	Lumen	% Total
0-10°	314 lm	2,9%
10-20°	908 lm	8,4%
20-30°	1431 lm	13,3%
30-40°	2008 lm	18,7%
40-50°	2710 lm	25,2%
50-60°	1979 lm	18,4%
60-70°	941 lm	8,7%
70-80°	343 lm	3,2%
80-90°	108 lm	1,0%
90-100°	8 lm	0,1%
100-110°	1 lm	0,0%
110-120°	1 lm	0,0%
120-130°	2 lm	0,0%
130-140°	2 lm	0,0%
140-150°	2 lm	0,0%
150-160°	1 lm	0,0%
160-170°	1 lm	0,0%
170-180°	0 lm	0,0%
<b>Total</b>	<b>10760 lm</b>	<b>100,0%</b>

### Intensity peaks

Max intensity	3602 cd
Intensity, 90°	33 cd
Intensity, 0°	3316 cd

### Zonal Lumen summary


Zone (γ)	Lumen	% Total
0-30°	2654 lm	24,7%
0-40°	4661 lm	43,3%
0-60°	9350 lm	86,9%
60-90°	1392 lm	12,9%
70-100°	459 lm	4,3%
90-120°	10 lm	0,1%
0-90°	10743 lm	99,8%
90-180°	17 lm	0,2%
0-180°	10760 lm	100,0%

### BUG rating

	Lumen	% Total
<b>Forward light</b>		
Low(0-30°)	1327 lm	12,3%
Medium(30-60°)	3341 lm	31,0%
High(60-80°)	648 lm	6,0%
Very high(80-90°)	55 lm	0,5%
<b>Back light</b>		
Low(0-30°)	1327 lm	12,3%
Medium(30-60°)	3341 lm	31,0%
High(60-80°)	648 lm	6,0%
Very high(80-90°)	55 lm	0,5%
<b>Uplight</b>		
Low(90-100°)	9 lm	0,1%
High(100-180°)	10 lm	0,1%

## IESNA TM-15-07 Luminaire Classification System For Outdoor Luminaires

**BUG rating B3 U1 G1**



# Light Measurement Report

Print date: 26-9-2025

Measurement date and time: 26-9-2025 10:48:27 – Measurement no. VFR-250926-3397-MS

Measurement tracking No. and Link: [VT250926-004750](#)

Operator:




## Power Details


### Input Power

Power feed to light source	76,8 W
Frequency of input power	50 Hz
RMS Input voltage feed, $V_{RMS}$	230 V
RMS Input current feed, $I_{RMS}$	0,344 A
Volt-Ampere or apparent power = $V_{RMS} * I_{RMS}$	79,11 VA
Displacement factor of AC power feed	0,97
Power factor of AC current feed	0,97
Total harmonic distortion of the current	4,23%
Total harmonic distortion of the voltage	0,07%

### Input Power Curve



### Efficiency



## Stabilization Details

### Warmup Conditions

Stable period	15 min
Stable change max	2,0%
Minimum time	15 min

### Color Temperature Change

CCT start	3986 K
CCT shift	+14 K
CCT end	4000 K


### Warmup Result

Total warmup time	Lamp stabilized in 15 min 2 sec
Warmup variation	+0,6%

### Output Change

Output start	10700 lm
Output change	+60 lm
Output end	10760 lm

### Stabilization Curve



# Light Measurement Report

Print date: 26-9-2025

Measurement date and time: 26-9-2025 10:48:27 – Measurement no. VFR-250926-3397-MS

Measurement tracking No. and Link: [VT250926-004750](https://www.viso-systems.com/VT250926-004750)

Operator:



## Flicker /TLA details

Flicker Meter Type: Viso Systems LabFlicker  
 Frequency of input power: 50 Hz  
 Flicker/TLA sample rate: 20000 samples/s

**Measurement time**  
 PstLM: 180 sec  
 All other indices: 1,2 sec

### Flicker indices according to Illuminating Engineering Society (IES)

Flicker frequency: 500 Hz  
 Percent Flicker: 0,36 %  
 Flicker index: 0

### Flicker indices according to California Energy Commission (CEC) 2016b

JA8/10 40 Hz: 0,22 %  
 JA8/10 90 Hz: 0,22 %  
 JA8/10 200 Hz: 0,26 %  
 JA8/10 400 Hz: 0,27 %  
 JA8/10 1000 Hz: 0,36 %


### TLA indices (re IEC TR 61547-1, IEC 61000-3-3 and IEC 61000-4-15)

PstLM value (F < 80 Hz): 0,1  
 SVM value (80 < F < 2000 Hz): 0

### Flicker indices according to Lighting Research Center (2015)

Perception metric, Assist Mp: 0,04

### Flicker frame (frame of one flicker period in time domain)



### Flicker FFT (flicker curve in frequency domain)



### IEEE 1789 Frequency/modulation plot

